

## **Clinical Investigation**

# **The Usefulness of Virtual Fluoroscopic Preprocedural Planning During Percutaneous Transhepatic Biliary Drainage**

Mitsuhiro Kinoshita<sup>1</sup>, Ryozo Shirono<sup>2</sup>, Katsuya Takechi<sup>2</sup>, Hironobu Yonekura<sup>3</sup>, Seiji Iwamoto<sup>4</sup>, Takayoshi Shinya<sup>1</sup>, Shoichiro Takao<sup>5</sup>, Masafumi Harada<sup>4</sup>

<sup>1</sup> Department of Radiology (Diagnostic Radiology), Tokushima University Hospital, 2-50-1, Kuramoto-cho, Tokushima City, Tokushima 770-8503, Japan

<sup>2</sup> Department of Radiology, Tokushima Red Cross Hospital, 103 Irinokuchi Komatsushima-cho, Komatsushima City, Tokushima 773-8502, Japan

<sup>3</sup> Department of Radiological Technology, Tokushima Red Cross Hospital, 103 Irinokuchi Komatsushima-cho, Komatsushima City, Tokushima 773-8502, Japan

<sup>4</sup> Department of Radiology and Radiation Oncology, Institute of Biomedical Sciences, Tokushima University Graduate School, 3-18-15, Kuramoto-cho, Tokushima City, Tokushima 770-8509, Japan

<sup>5</sup> Department of Diagnostic Radiology, Tokushima University Graduate School of Health Science, 3-18-15, Kuramoto-cho, Tokushima City, Tokushima 770-8509, Japan

Corresponding author: Mitsuhiro Kinoshita

Department of Radiology (Diagnostic Radiology), Tokushima University Hospital,

2-50-1, Kuramoto-cho, Tokushima City, Tokushima 770-8503, Japan

Tel: +81-88-633-7173

Fax: +81-88-633-7468

E-mail: kinoshita.3216@tokushima-u.ac.jp

**Acknowledgments:** None.

**Funding:** None.

**Conflict of interest:** The authors have no conflicts of interest to declare.

1 **Abstract**

2 Purpose: To retrospectively evaluate the usefulness of virtual fluoroscopic preprocedural  
3 planning (VFPP) in the percutaneous transhepatic biliary drainage (PTBD) procedure.

4 Materials and Methods: Twenty-two patients who were treated by PTBD were included  
5 in this study. Twelve patients were treated using PTBD intraoperative referencing  
6 coronal computed tomography (CT) images (i.e., coronal CT group), and 10 patients  
7 were treated using PTBD intraoperative referencing VFPP images (i.e., VFPP group).  
8 To analyze the effect of the intraoperative referencing VFPP image, the VFPP group  
9 was retrospectively compared with the coronal CT group.

10 Results: The characteristics of both patient groups were not statistically significantly  
11 different. There were no significant differences in the targeted bile duct, diameter and  
12 depth of the target bile, breath-holding ability, number of targeted bile duct puncture  
13 attempts, change in the targeted bile duct, and exchange of the drainage catheter.

14 However, the X-ray fluoroscopy time and the procedure time were significantly shorter  
15 in the VFPP group than in the coronal CT group (196 vs. 334 seconds,  $P < 0.05$ ; and  
16 16.0 vs. 27.2 minutes,  $P < 0.05$ ).

17 Conclusion: Intraoperative referencing using the VFPP imaging in PTBD intuitively can  
18 be a useful tool for better localization of the guidewire in the bile duct, and thereby

1 shorten the X-ray fluoroscopy time and procedure time while minimizing radiation  
2 exposure and complications.  
3  
4 Keywords: Virtual fluoroscopic preprocedural planning • Virtual fluoroscopy •  
5 Percutaneous transhepatic biliary drainage • Procedure time • Ray Summation image •  
6 X-ray fluoroscopy time  
7  
8  
9  
10  
11  
12  
13  
14  
15  
16  
17  
18  
19  
20  
21

1 **Introduction**

2 Percutaneous transhepatic biliary drainage (PTBD) is an established interventional  
3 procedure for treating patients with malignant or benign biliary obstruction [1-7]. A  
4 recent quality improvement guideline for PTBD reported that the threshold of success  
5 rate for dilated ducts is 95% [5]. However, PTBD is an invasive technique and has some  
6 associated complications [6]. The recommended overall procedure threshold for all  
7 major complications of PTBD is 10% [5]. Complications can be categorized as “early”  
8 (i.e., occurring within 30 days) or “late”. Approximately one-half of early complications  
9 are procedure-related [8-10]. Thus, it is important that the procedure is performed  
10 precisely and quickly to reduce complications.

11 We have used virtual fluoroscopic preprocedural planning (VFPP) as a new  
12 navigation system to perform PTBD precisely and quickly. Virtual fluoroscopic  
13 preprocedural planning involves using an image in which trace lines are drawn along  
14 ductal structures (e.g., vessels or the biliary duct) on a Ray Summation image, which is  
15 similar to a fluoroscopy image, and it is easy to prepare using volume data from  
16 multiple-detector computed tomography (MDCT) and workstation [11]. Fukusumi et al.  
17 [11] reported the feasibility of VFPP in bronchoscopy. However, to our knowledge,  
18 there has yet been no report on the usefulness of VFPP in the PTBD procedure.

1 Therefore, the purpose of this study was to retrospectively evaluate the usefulness of  
2 VFPP in the PTBD procedure.

3

#### 4 **Materials and Methods**

##### 5 Patients

6 Between April 2014 and April 2016 in the Department of Radiology at a single  
7 hospital, first-attempt PTBD was performed in 24 consecutive patients with obstructive  
8 jaundice due to biliary obstruction. We excluded 2 patients because the first-attempt  
9 PTBD was unsuccessful on account of the small caliber of the target duct. Thus, 22  
10 patients (11 men and 11 women; mean age, 76.8 years; range, 62–91 years) were  
11 included in this study. The causes of malignant biliary obstruction were pancreatic  
12 cancer (n = 7), cholangiocarcinoma (n = 3), lymph node metastasis of gastric cancer (n  
13 = 3), recurrent gastric cancer (n = 2), carcinoma of the ampulla of Vater (n = 1),  
14 gallbladder cancer (n = 1), and right renal pelvic cancer (n = 1). The cause of benign  
15 biliary obstruction was choledocholithiasis (n = 4). Twelve patients were treated by  
16 PTBD using intraoperative referencing coronal computed tomography (CT) images (i.e.,  
17 coronal CT group) and 10 patients were treated by PTBD using intraoperative VFPP  
18 images (i.e., VFPP group). The institutional review board approved this retrospective

1 study and no individual patient consent was required. All patients were informed about  
2 the benefits and potential risks of the procedure. All patients provided written informed  
3 consent.

4

#### 5 The PTBD Procedure

6 All procedures were performed by 2 interventional radiologists with at least 10  
7 years of experience with interventional techniques in the angiography suite. All patients  
8 were initially scanned by ultrasound (US) using a 3.5 MHz electronic convex probe  
9 (F37; Hitachi, Tokyo, Japan), and a CT scan was performed using the Somatom  
10 Sensation Cardiac 64 scanner (Siemens Medical Solutions, Forchheim, Germany). The  
11 PTBD procedures were administered in standard fashion, as described elsewhere [2, 3,  
12 10].

13 After administering local anesthesia with 1% lidocaine (Xylocaine Injection;  
14 AstraZeneca, Osaka, Japan) and intravenous sedation with pentazocine hydrochloride  
15 (Sosegon Injection; Maruishi Pharmaceutical, Osaka, Japan), a 18-gauge needle (Create  
16 Medic, Kanagawa, Japan) was inserted under US guidance using a 3.5MHz  
17 intraoperative electronic convex probe (F37; Hitachi, Tokyo, Japan) with a puncture  
18 adapter. Once the backflow of bile occurred and on successfully placing the needle tip

1 in the bile duct, a 0.035-inch guidewire (Radifocus; Terumo, Tokyo, Japan) was  
2 advanced into the proximal obstruction under X-ray fluoroscopy guidance while using  
3 the reference coronal CT images or VFPP images. After removing the needle, a  
4 7-French dilator (Cook Japan, Tokyo, Japan) and 7.2-French pigtail drainage catheter  
5 (Cook Japan, Tokyo, Japan) were inserted in order under X-ray fluoroscopy guidance.  
6 After placing the drainage catheter, a small amount of dilute contrast media was infused  
7 through the catheter to confirm that the catheter was in the appropriate position. If  
8 puncturing the targeted bile duct proved difficult, despite several attempts, then the  
9 other duct was punctured, and the change in protocol was recorded. If a guidewire could  
10 be advanced beyond the obstruction of the common bile duct, a 7.2-French pigtail  
11 drainage catheter, which was customized by adding side holes at the appropriate  
12 position (internal-external catheter), was placed into the small bowel.

13

#### 14 Coronal CT Imaging

15 For all patients, MDCT was performed by using a Somatom Sensation Cardiac 64  
16 scanner. The range of CT covered the area of the abdomen. The scan and reconstruction  
17 parameters of CT were  $64 \times 0.6$  mm beam collimation, 120 kV X-ray tube voltage, 250  
18 reference mAs of the X-ray tube current, and automatic exposure control (CARE Dose

1 4D; Siemens Medical Systems, Forchheim, Germany), 0.5-sec gantry rotation time,  
2 pitch factor of 1.1, 3.0 mm slice thickness, and reconstruction intervals of 3.0 mm. The  
3 convolution kernel was B31f medium smooth+.

4 For the protocol with contrast material (CM), 600 mg iodine per kilogram body  
5 weight were injected for 30 seconds ( $20 \text{ [mg} \cdot \text{kg}^{-1}\text{]/sec}$  fractional dose) and the CT scan  
6 delay times were 45 seconds and 120 seconds. The CM was iopamidol 300 or iopamidol  
7 370 (Iopamiron with 300 mg or 370 mg of iodine per milliliter; Bayer Pharma, Osaka,  
8 Japan), depending on the patient's weight. It was injected into the brachial vein with a  
9 22-gauge needle at a rate of 2.1–3.2 mL/sec using an automated injector (Dual Shot;  
10 Nemoto Kyorindo, Tokyo, Japan).

11 Contrast-enhanced MDCT was performed to all patients with a malignant biliary  
12 obstruction ( $n = 18$ ). However, plain MDCT was performed to all patients with a benign  
13 biliary obstruction ( $n = 4$ ).

14

#### 15 Virtual Fluoroscopic Preprocedural Planning

16 Since April 2015, we have used VFPP as a new navigation system. Virtual  
17 fluoroscopic preprocedural planning uses an image that is created from the volume data  
18 obtained from MDCT imaging with a workstation; the trace lines of the bile duct or

1 vessels on the Ray Summation image are similar to those of a fluoroscopy image [6].  
2 Multidetector CT images of 1.0 mm thickness were transferred to a 3-dimensional  
3 workstation (SynapseVincent; Fujifilm Co., Tokyo, Japan). By referring to multiplanar  
4 reconstruction CT images, we drew a trace line between the targeted bile duct and the  
5 common bile duct before it enters the duodenum at the major duodenal papilla along the  
6 center of the bile duct. In addition, if necessary, we drew trace lines along neighboring  
7 bile ducts and added annotations at the obstruction and the major duodenal papilla. The  
8 lines could be displayed at any angle with 3-dimensional virtual fluoroscopic image. We  
9 reconstructed the background by using a Ray Summation image, and applied an edge  
10 filter (Figure 1). Customization of using workstation made it possible to create this  
11 image in approximately 10 minutes. All VFPPs were performed by 2 authors, and  
12 included a radiologist and radiological technician. In instances of disagreement, a final  
13 consensus was reached through interobserver discussion. We applied VFPP to the PTBD  
14 procedure. Virtual fluoroscopic preprocedural planning was used as a reference to select  
15 the bile duct during the PTBD procedure.

16

## 17 Assessment

18 To analyze the effect of the intraoperative referencing VFPP image, 10 consecutive

1 patients who had undergone first-attempt PTBD using VFPP (i.e., VFPP group) were  
2 retrospectively compared with 12 consecutive patients who had undergone first-attempt  
3 PTBD using coronal CT images (i.e., coronal CT group). The Fisher's exact test or  
4 Mann–Whitney *U* test was used to evaluate significant differences in the patients'  
5 characteristics and results of the PTBD procedure in the coronal CT and VFPP groups.  
6 Categorical variables are expressed as the number and compared by Fisher's exact test.  
7 Continuous variables are expressed as the mean  $\pm$  the standard deviation and  
8 compared by the Mann–Whitney *U* test. The patients' characteristics that were  
9 examined were age, sex, body weight, etiology (i.e., benign or malignant biliary  
10 obstruction), location of the obstructing lesion (i.e., upper or lower), ascites, liver  
11 toxicity, inflammation, and jaundice. We classified the location of the obstructing lesion  
12 into 2 categories, based on the intraoperative cholangiopancreatography findings: (1)  
13 the “upper” location indicated that the obstructing lesion was above the point of the  
14 communication of the 3 ducts (i.e., cystic, common hepatic, and common bile duct) and  
15 (2) the “lower” location indicated that the obstructing lesion was below the point of the  
16 communication of the 3 ducts. Liver toxicity was evaluated by the aspartate  
17 aminotransferase (AST) and alanine aminotransferase (ALT) levels; inflammation was  
18 evaluated by white blood cell counts and C-reactive protein levels; and jaundice was

1 evaluated by the serum total bilirubin level by laboratory tests that were performed  
2 within 2 days before the PTBD procedure. Percutaneous transhepatic biliary drainage  
3 involved the targeted bile duct, diameter of the bile duct, depth of the bile duct,  
4 breath-holding ability (i.e., good or poor), number of targeted bile duct puncture  
5 attempts, change in the targeted bile duct, exchange of the drainage catheter, X-ray  
6 fluoroscopy time, and procedure time. We classified breath-holding into 2 grades,  
7 according to our instructions: (1) “good” was defined as breath-holding for more than 3  
8 seconds and (2) “poor” was defined as breath-holding for less than 3 seconds. We  
9 defined the procedure time as the time from the administration of the local anesthesia  
10 until confirming, using the contrasting media, that the catheter was in the appropriate  
11 position. Statistical significance was based on 2-sided design-based tests evaluated at  $\alpha$   
12 = 0.05. All statistical analyses were performed using SPSS, version 23 (IBM, Armonk,  
13 NY, USA).

14

## 15 **Results**

### 16 The Patients’ Characteristics

17 The characteristics of both patient groups are presented in Table 1. There were no  
18 statistically significant differences between the groups.

1

## 2 Results of the PTBD Procedure

3 The results of PTBD procedure of both groups are presented in Table 2. The targeted  
4 bile duct was B3 or B8. In the coronal CT group, B3 was targeted in 8 patients and B8  
5 was targeted in 4 patients. In the VFPP group, B3 was targeted in 9 patients and B8 was  
6 targeted in one patient. There were no statistically significant differences between the  
7 groups.

8 It was necessary to change the targeted bile duct in 2 patients in the coronal CT  
9 group and in one patient in the VFPP group because we failed to puncture the bile duct.  
10 There were no statistically significant differences between the groups.

11 There were also no significant differences in the targeted bile duct, diameter and  
12 depth of target bile, breath-holding, number of puncture attempts, change in the targeted  
13 bile duct, and exchange of the drainage catheter. The X-ray fluoroscopy time and the  
14 procedure time were significantly shorter in the VFPP group than in the coronal CT  
15 group (196 seconds vs. 334 seconds,  $P < 0.05$ ; and 16.0 minutes vs. 27.2 minutes,  $P <$   
16 0.05).

17 There were 6 procedure-related minor complications (Society of Interventional  
18 Radiology [SIR] Classification B). One patient in each group developed mild pain at the

1 puncture site and 2 patients in each group developed mild abdominal pain. One  
2 procedure-related major complication (SIR Classification D) occurred. One patient in  
3 the coronal CT group experienced subcapsular hepatic hematoma, but this patient was  
4 successfully treated conservatively.

5

## 6 **Discussion**

7 The importance of using imaging in preprocedural planning for patients with bile  
8 duct obstruction cannot be overemphasized because the level of obstruction and bile  
9 duct anatomy are most easily determined by MDCT or by magnetic resonance  
10 cholangiopancreatography [7]. However, it is impossible by MDCT to display the bile  
11 duct anatomy and obstruction in one image. Magnetic resonance  
12 cholangiopancreatography displays the bile duct anatomy and obstruction in one image,  
13 but the image is not similar to a fluoroscopy image. By contrast, a VFPP image displays  
14 the bile duct anatomy and obstruction in only one image that is similar to a fluoroscopy  
15 image (Figure 2). Because the lines can be displayed at any angle in the 3-dimensional  
16 virtual fluoroscopic image, we could also choose an appropriate working angle  
17 beforehand and perform the procedure in this state (Figure 3). Moreover, VFPP images  
18 are effective in the situation of a targeted bile duct with a bending and meandering

1 course.

2 Using intraoperative referencing VFPP images enables a physician to advance the  
3 guidewire precisely and quickly into the common bile duct under X-ray fluoroscopy  
4 guidance. Therefore, it was possible to shorten the X-ray fluoroscopy time and  
5 procedure time by using VFPP. Shortening the X-ray fluoroscopy time may contribute  
6 to lower amounts of radiation exposure to the patient and staff.

7 One-half of early complications are procedure-related [8-10]. These complications  
8 include pain at the puncture site, bile leak with the risk of biliary peritonitis and biloma  
9 formation, cholangitis, septicemia, and bleeding [9, 10]. Adequate preprocedural  
10 planning may help to reduce these complications.

11 New navigation systems are used in various procedures. The feasibility of  
12 automated tumor-feeders detection software for hepatocellular carcinoma and  
13 fluoroscopy with magnetic resonance angiography/computerized tomographic  
14 angiography fusion guidance for peripheral artery interventions have been reported in  
15 the literature [12-14]. These navigation systems are quite useful, but they are not  
16 available at all medical institutions. By contrast, VFPP is easy to prepare if only the  
17 volume data from MDCT and a workstation are available [11]. The introduction of a  
18 new device is therefore unnecessary, and an extra financial burden is avoided.

1 In this study, we were able to create VFPP from contrast-enhanced MDCT in all  
2 patients, except for one patient. All bile ducts were dilated. However, VFPP may take a  
3 long time if only plain MDCT images are available or if the bile duct does not dilate.

4 In this study, the X-ray fluoroscopy time was 73–783 seconds (mean time,  
5 271seconds). Kloeckner et al. [15] reported that the mean X-ray fluoroscopy times of  
6 right- and left-sided PTBD were 13.5 minutes and 16.6 minutes. The mean PTBD time  
7 was longer in that study than in our study. There are 2 reasons for this finding. First, we  
8 performed PTBD in a one-step method using an 18-gauge needle instead of in the  
9 two-step method because all patients in this study had dilated ducts [10]. Second, in  
10 only 2 patients, we placed the internal-external catheter beyond the obstruction of the  
11 common hepatic duct, and we placed the obligatory external drainage catheter above the  
12 obstruction in many patients.

13 This study had several limitations. First, it was a single-center study. Second, the  
14 retrospective nature of the study may also limit the conclusions drawn from it. Third,  
15 the number of patients in the VFPP group and CT group was uneven, which may have  
16 introduced bias in the results. However, the baseline characteristics of patients in both  
17 groups were equivalent. Fourth, the accuracy of VFPP depends on the planner because  
18 the lines in VFPP are drawn manually. However, all VFPP procedures were performed

1 by 2 authors, and included a radiologist and radiological technician. It was believed that  
2 the VFPP images were sufficient to be used as reference images.

3

#### 4 **Conclusion**

5 Intraoperative referencing of the VFPP image during PTBD can intuitively be a  
6 useful tool for better localization of the guidewire in the bile duct. This procedure  
7 thereby shortens the X-ray fluoroscopy time and procedure time while minimizing  
8 radiation exposure and complications.

9

10

11

## 1   **References**

- 2   1. Takada T, Hanyu F, Kobayashi S, Uchida Y. Percutaneous transhepatic cholangial  
3       drainage: direct approach under fluoroscopic control. *J Surg Oncol.* 1976;8:83–97.
- 4   2. Makuuchi M, Bandai Y, Ito T, Watanabe G, Wada T, Abe H, Muroi T. Ultrasonically  
5       percutaneous transhepatic bile drainage: a single-step procedure without  
6       cholangiography. *Radiology.* 1980;136:165–9.
- 7   3. Baxter-Smith DC, Temple JG, Howarth F. Palliative percutaneous transhepatic  
8       drainage for inoperable obstructive jaundice. *Ann R Coll Surg Engl.* 1982;64:394–6.
- 9   4. Nagino M, Hayakawa N, Nimura Y, Dohke M, Kitagawa S. Percutaneous  
10       transhepatic biliary drainage in patients with malignant biliary obstruction of the  
11       hepatic confluence. *Hepatogastroenterology.* 1992;39:296–300.
- 12   5. Saad WE, Wallace MJ, Wojak JC, Kundu S, Cardella JF. Quality improvement  
13       guidelines for percutaneous transhepatic cholangiography, biliary drainage, and  
14       percutaneous cholecystostomy. *J Vasc Interv Radiol.* 2010;21:789–95.
- 15   6. Hamlin J, Friedman M, Stein M, Bray J. Percutaneous biliary drainage:  
16       complications of 118 consecutive catheterizations. *Radiology.* 1986;158:199–202.
- 17   7. Covey AM, Brown KT. Percutaneous transhepatic biliary drainage. *Tech Vasc Interv*  
18       *Radiol.* 2008;11:14–20.

- 1 8. Saluja SS, Gulati M, Garg PK, Pal H, Pal S, Sahni P, Chattopadhyay TK.  
2 Endoscopic or percutaneous biliary drainage for gallbladder cancer: a randomized  
3 trial and quality of life assessment. *Clin Gastroenterol Hepatol*. 2008;6:944–50.
- 4 9. Walter T, Ho CS, Horgan AM, Warkentin A, Gallinger S, Greig PD, Kortan P, Knox  
5 JJ. Endoscopic or percutaneous biliary drainage for Klatskin tumors? *J Vasc Interv  
6 Radiol*. 2013;24:113–21.
- 7 10. Madhusudhan KS, Gamanagatti S, Srivastava DN, Gupta AK. Radiological  
8 interventions in malignant biliary obstruction. *World J Radiol*. 2016;8:518–29.
- 9 11. Fukusumi M, Ichinose Y, Arimoto Y, Takeoka S, Homma C, Matsuoka H, Mouri A,  
10 Hamamoto Y, Matsumoto J, Kamimura M. Bronchoscopy for pulmonary peripheral  
11 lesions with virtual fluoroscopic preprocedural planning combined with EBUS-GS:  
12 a pilot study. *J Bronchology Interv Pulmonol*. 2016;23:92–7.
- 13 12. Miyayama S, Yamashiro M, Hashimoto M, Hashimoto N, Ikuno M, Okumura K,  
14 Yoshida M, Matsui O. Identification of small hepatocellular carcinoma and  
15 tumor-feeding branches with cone-beam CT guidance technology during  
16 transcatheter arterial chemoembolization. *J Vasc Interv Radiol*. 2013; 24:501–8.
- 17 13. Miyayama S, Yamashiro M, Ikuno M, Okumura K, Yoshida M. Ultraselective  
18 transcatheter arterial chemoembolization for small hepatocellular carcinoma guided

- 1 by automated tumor-feeders detection software: technical success and short-term  
2 tumor response. *Abdom Imaging*. 2014;39:645–56.
- 3 14. Sailer AM, de Haan MW, de Graaf R, van Zwam WH, Schurink GWH, Nelemans PJ,  
4 Wildberger JE, Das M. Fusion guidance in endovascular peripheral artery  
5 interventions: a feasibility study. *Cardiovasc Intervent Radiol*. 2015;38:314–21.
- 6 15. Kloeckner R, Bersch A, Dos Santos DP, Schneider J, Düber C, Pitton MB. Radiation  
7 exposure in nonvascular fluoroscopy-guided interventional procedures. *Cardiovasc*  
8 *Intervent Radiol*. 2012;35:613–20.

## FIGURE LEGENDS

**Fig. 1** The process of VFPP using a workstation.

**a** and **b** While referencing the multiplanar reconstruction computed tomography images, we drew a trace line between the targeted bile duct, neighboring bile ducts, and common bile duct before it enters the duodenum at the major duodenal papilla along the center of the bile duct. **c** The background is a reconstructed Ray Summation image. **d** An edge filter is applied to the background

VFPP, virtual fluoroscopic preprocedural planning

**Fig. 2** An 85-year-old female with obstructive jaundice due to pancreatic cancer.

**a** We used this VFPP image as the reference image. The yellow line shows the course of B3 to the common bile duct. The annotation is the obstruction in the common bile duct. The blue lines show neighboring bile ducts. The VFPP image is similar to the fluoroscopy image. **b** The fluoroscopy image was obtained by infusing a small amount of dilute contrast media through the catheter

VFPP, virtual fluoroscopic preprocedural planning

**Fig. 3** The VFPP makes it possible to display the three-dimensional virtual fluoroscopic

image at any angle.

**a** The frontal view of B3 to the common bile duct. **b** The right anterior view (RAO) 25° of B3 to the common bile duct. **c** The left anterior view (LAO) 25° of B3 to the common bile duct

VFPP, virtual fluoroscopic preprocedural planning

**Table 1** The patients' characteristics

	Coronal CT group ( <i>n</i> = 12)	VFPP group ( <i>n</i> = 10)	<i>P</i> value
Age (year)	74.6 ± 7.29	79.4 ± 9.67	0.159
Sex (male/female)	8/4	3/7	0.198
Body weight (kg)	43.5 ± 11.8	48.2 ± 7.08	0.123
Etiology (benign/malignant)	3/9	1/9	0.594
Location of the obstructing lesion (upper/lower)	0/12	2/8	0.195
Ascites (+/-)	4/8	2/8	0.646
Preoperative TB (mg/dL)	5.3 ± 3.4	6.2 ± 5.0	0.628
Preoperative AST (U/L)	165 ± 103	172 ± 106	0.872
Preoperative ALT (U/L)	146 ± 121	169 ± 136	0.722
Preoperative WBC (/μL)	9778 ± 6289	10180 ± 8809	0.872
Preoperative CRP (mg/dL)	7.00 ± 8.30	6.41 ± 6.50	0.923

*ALT* alanine aminotransferase, *AST* aspartate aminotransferase, *CRP* C-reactive protein, *CT* computed tomography, *PTBD* percutaneous transhepatic biliary drainage, *TB* total bilirubin, *VFPP* virtual fluoroscopic preprocedural planning, *WBC* white blood cell count


**Table 2** Results of the PTBD procedure

	Coronal CT group ( <i>n</i> = 12)	VFPP group ( <i>n</i> = 10)	<i>P</i> value
Targeted bile duct (B3/B8)	8/4	9/1	0.323
Diameter of the bile duct (mm)	7.1 ± 2.8	7.2 ± 2.7	0.872
Depth of the bile duct (mm)	40.9 ± 12.1	31.9 ± 11.0	0.228
Breath-holding ability (good/poor)	11/1	9/1	1.000
Number of targeted puncture attempts (time)	1.7 ± 1.0	1.4 ± 1.0	0.456
Change in the targeted bile duct (yes/no)	2/10	1/9	1.000
Exchange of the drainage catheter (yes/no)	1/11	1/9	1.000
X-ray fluoroscopy time (sec)	334 ± 231	196 ± 137	0.043*
Procedure time (min)	27.2 ± 13.9	16.0 ± 9.65	0.021*



\* Indicates a significant difference

*CT* computed tomography, *PTBD* percutaneous transhepatic biliary drainage, *VFPP* virtual fluoroscopic preprocedural planning

Fig. 1



**Fig. 2**



**Fig. 3**

